

Provided By:

General Air Service & Supply
6330 Colorado Blvd
Commerce City, CO 80022
Phone: (303) 892-7003
www.generalair.com

PRODUCT Carbon Dioxide

CGA CONNECTION 320

GRADE	PURITY	Air	THC	H ₂ O	N ₂	O ₂	CH ₄	H ₂ S	NO/NO ₂	SO ₂	CO	ID
Instrument	99.99%	-	10 PPM	10 PPM	70 PPM	15 PPM	-	-	-	-	-	-
Coleman	99.99%	-	-	10 PPM	-	20 PPM	-	-	-	-	-	-
Bone Dry	99.90%	-	-	10 PPM	-	-	-	-	-	-	-	-
Lab	99.00%	1.00%	-	30 PPM	-	-	-	-	-	-	-	-
Beverage	99.90%	-	20 PPM	20 PPM	-	30 PPM	-	-	-	-	-	-
Industrial	98.00%	-	10 PPM	120 PPM	-	-	-	-	-	-	-	-
USP	99.00%	1.00%	-	150 mg/m ³	-	-	25 PPM	1 PPM	2.5 PPM	5 PPM	10 PPM	PASS

The above information represents the specifications for the denoted product. The actual gas content may vary and can only be verified by an individual cylinder analysis.